

AUG 23 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130306</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>C. RITAK</u>	ID# <u>3791</u> ✓ Date <u>8/23/10</u> ✓
A Agency <u>APD</u>	Phone # <u>786 2640</u>
Instrument Location <u>4th Ave Substation (FLORA)</u>	
Alco S/N <u>B72446</u> ✓ Target Value <u>.082</u> ✓ High Pressure <u>1150</u>	
B Alco Test Values <u>0.00</u> ✓ <u>0.03</u> ✓	
1st Alco 2nd Alco	
Signature <u>[Signature]</u> <u>RITAK/1349</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of SEPT, 2010.

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130306

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

AUGUST 23, 2010

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

SIXTY DAY

O.L. #: A

DEPT/AGENCY: ANC1

CASE/REPORT: A

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: B72446

— BREATH ANALYSIS —

.002 ADJUSTED FOR 30.16 in		
ALCO TARGET	.002	23:46
BLANK TEST	.000	23:47
INTERNAL STANDARD	VERIFIED	23:47
ALCO TV 30.16 in	.000	23:48
BLANK TEST	.000	23:48
SUBJECT SAMPLE	.000	23:49
BLANK TEST	.000	23:50
ALCO TV 30.16 in	.003	23:50
BLANK TEST	.000	23:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

AUGUST 23, 2010

TIME 23:44

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 30.16 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢